

Code No. 27102

**Mouse VEGF Assay Kit - IBL****INTRODUCTION**

Vascular Endothelial Cell Growth Factor (VEGF) is a homodimeric protein initially purified from media conditioned by normal bovine pituitary folliculo-stellate cells and secreted by a variety of vascularized tissues. It was subsequently found to be identical to a vascular permeability factor (VPF), which was previously identified in media conditioned by tumor cell lines based upon its ability to increase the permeability of capillary blood vessels. The reported activities of VEGF include stimulation of endothelial cell growth, angiogenesis and capillary permeability. In normal tissues, VEGF expression has been observed in activated macrophages, keratinocytes, hepatocytes, smooth muscle cells Leydig cells, embryonic fibroblasts and bronchial and choroid plexus epithelium, renal glomerular visceral epithelium and mesangial cells.

**PRINCIPLE**

This kit is a solid phase sandwich ELISA using 2 kinds of high specific antibodies. Tetra Methyl Benzidine (TMB) is used as a coloring agent (Chromogen). The strength of coloring is proportional to the quantities of Mouse VEGF.

**MEASUREMENT RANGE**

62.5 – 4,000 ng/mL

**INTENDED USE****For research use only, not for use in diagnostic procedures.**

- This IBL's assay kit is capable for the quantitative determination of Mouse VEGF in EDTA-plasma, serum and supernatant of cell culture media.
- Both recombinant and native form of Mouse VEGF can be detected with this kit.

**KIT COMPONENT**

1	Precoated plate : Anti-Mouse VEGF (V-N) Rabbit IgG Affinity Purify	96Well x 1
2	Labeled antibody Conc. : (30X) HRP conjugated Anti- Mouse VEGF Rabbit IgG Fab' Affinity Purify	0.4mL x 1
3	Standard : Recombinant Mouse VEGF 164	0.5mL x 2
4	EIA buffer *	30mL x 1
5	Solution for Labeled antibody *	12mL x 1
6	Chromoge : TMB solution	15mL x 1
7	Stop solution *	12mL x 1
8	Wash buffer Conc. *	50mL x 1

**OPERATION MANUAL****1. Materials needed but not supplied**

- Plate reader (450nm)
- Graduated cylinder and beaker
- Refrigerator (as 4°C)
- Paper towel
- Incubator (37°C ± 1°C)
- Washing bottle for precoated plate
- Disposable test tube for "2, Labeled antibody Conc." and "6, Chromogen"
- Micropipette and tip
- Deionized water
- Graph paper (log/log)
- Tube for dilution of Standard


**2. Preparation**

- 1) Preparation of wash buffer  
"8, Wash buffer Conc." is a concentrated (40X) buffer. Adjust the temperature of "8, Washing buffer Conc." to room temperature and then, mix it gently and completely before use. Dilute 50 mL of "8, Wash buffer Conc." with 1,950 mL of deionized water and mix it. This is the wash buffer for use. This prepared wash buffer shall be stored in refrigerator and used within 2 weeks after dilution.
- 2) Preparation of Labeled antibody  
"2, Labeled antibody Conc." is a concentrated (30X). Dilute "2, Labeled antibody Conc." with "5, Solution for Labeled antibody" in 30 times according to required quantity into a disposable test tube. Use this resulting solution as Labeled antibody.  
Example)  
In case you use one strip (8 well), the required quantity of Labeled antibody is 800 µL. (Dilute 30 µL of "2, Labeled antibody Conc." with 870 µL of "5, Solution for Labeled antibody" and mix it. And use the resulting solution by 100 µL in each well.)  
This operation should be done just before the application of Labeled antibody.  
The remaining "2, Labeled antibody Conc." should be stored at 4°C in firmly sealed vial.
- 3) Preparation of Standard  
Put just 0.5 mL of deionized water into the vial of "3, Standard" and mix it gently and completely. This solution is 8,000 pg/mL Mouse VEGF standard.
- 4) Dilution of Standard  
Prepare 8 tubes for dilution of "3, Standard". Put 230 µL each of "4, EIA buffer" into the tube.  
Specify the followipg concentration of each tube."  

Tube-1	4,000 pg/mL
Tube-2	2,000 pg/mL
Tube-3	1,000 pg/mL
Tube-4	500 pg/mL
Tube-5	250 pg/mL
Tube-6	125 pg/mL
Tube-7	62.5 pg/mL
Tube-8	0 pg/mL (Test Sample Blank)

Put 230 µL of Standard solution into tube-1 and mix it gently. Then, put 230 µL of tube-1 mixture into tube-2. Dilute two times standard solution in series to set up 7 points of diluted standard between 4,000 pg/mL and 62.5 pg/mL. Tube-8 is the test sample blank as 0 pg/mL.

See followipg picture.

**5) Dilution of test sample**

Test samples should be diluted with "4, EIA buffer" as necessary.

If the concentration of Mouse VEGF in samples may not be estimated in advance, the pre-assay with several different dilutions will be recommended to determine the proper dilution of samples.

**3. Measurement procedure**

All reagents shall be brought to room temperature approximately 30 minutes before use. Then mix it gently and completely before use. Make sure of no change in quality of the reagents. Standard curve shall be prepared simultaneously with the measurement of test samples.

Reagents	Test Sample	Standard	Test Sample Blank	Reagent Blank
	Test sample 100 µL	Diluted standard (Tube 1-7) 100 µL	EIA buffer (Tube-8) 100 µL	EIA buffer 100 µL
Incubation for 60 minutes at 37°C with plate lid				
4 times (wash buffer more than 350 µL)*				
Labeled Antibody	100 µL	100 µL	100 µL	-
Incubation for 30 minutes at 4°C with plate lid				
5 times (wash buffer more than 350 µL)*				
Chromogen	100 µL	100 µL	100 µL	100 µL
Incubation for 30 minutes at room temperature (shielded)				
Stop solution	100 µL	100 µL	100 µL	100 µL
Read the plate at 450nm against a Reagent Blank within 30 minutes after addition of Stop solution.				

- 1) Determine wells for reagent blank. Put 100 µL each of "4, EIA buffer" into the wells.
- 2) Determine wells for test sample blank, test sample and diluted standard. Then, put 100 µL each of test sample blank (tube-8), test sample and dilutions of standard (tube-1-7) into the appropriate wells.
- 3) Incubate the precoated plate for 60 minutes at 37°C after covering it with plate lid.
- 4) Wash the plate with the prepared wash buffer and remove all liquid.\*
- 5) Pipette 100 µL of labeled antibody solution into the wells of test samples, diluted standard and test sample blank.
- 6) Incubate the precoated plate for 30 minutes at 4°C after covering it with plate lid.
- 7) Wash the plate with the prepared wash buffer and remove all liquid.\*
- 8) Take the required quantity of "6, Chromogen" into a disposable test tube. Then, pipette 100 µL from the test tube into the wells. Please do not return the rest of the test tube to "6, Chromogen" bottle to avoid contamination.
- 9) Incubate the precoated plate for 30 minutes at room temperature in the dark. The liquid will turn blue by addition of "6, Chromogen".
- 10) Pipette 100 µL of "7, Stop solution" into the wells. Mix the liquid by tapping the side of precoated plate. The liquid will turn yellow by addition of "7, Stop solution".
- 11) Remove any dirt or drop of water on the bottom of the precoated plate and confirm there is no bubble on the surface of the liquid. Then, run the plate reader and conduct measurement at 450 nm against a reagent blank. The measurement shall be done within 30 minutes after addition of "7, Stop solution".

**SPECIAL ATTENTION**


- 1) Test samples should be measured soon after collection. For the storage of test samples, store them frozen and do not repeat freeze/thaw cycles. Thaw the test samples at a low temperature and mix them completely before measurement.
- 2) Test samples should be diluted with "4, EIA buffer", if the need arises.
- 3) Duplicate measurement of test samples and standard is recommended.
- 4) Use test samples in neutral pH range. The contaminations of organic solvent may affect the measurement.
- 5) Use only wash buffer contained in this kit for washing the precoated plate. Insufficient washing may lead to the failure in measurement.
- 6) Remove the wash buffer completely by tapping the precoated plate on paper towel. Do not wipe wells with paper towel.
- 7) "6, Chromogen" should be stored in the dark due to its sensitivity against light. "6, Chromogen" should be avoided contact with metals.
- 8) Measurement should be done within 30 minutes after addition of "7, Stop solution".

**CALCULATION OF TEST RESULT**

Subtract the absorbance of test sample blank from all data, including standards and unknown samples before plotting. Plot the subtracted absorbance of the standards against the standard concentration on log-log graph paper. Draw the best smooth curve through these points to construct the standard curve. Read the concentration for unknown samples from the standard curve.

**Example of standard curve**

Conc. (pg/mL)	Absorbance (450nm)
4,000	2.089
2,000	1.238
1,000	0.670
500	0.365
250	0.196
125	0.109
62.5	0.077
0 (Test Sample Blank)	0.020



\* The typical standard curve is shown above. This curve can not be used to derive test results. Please run a standard curve for each assay.

- your skin and clothes contact "7, Stop solution" and pay attention to the disposal of "7, Stop solution".
- Dispose used materials after rinsing them with large quantity of water.
- Precipitation may occur in "2, Labeled antibody Conc.", however, there is no problem in the performance.
- Wash hands after handling reagents.
- Do not mix the reagents with the reagents from a different lot or kit.
- Do not use expired reagents.
- This kit is for research purpose only. Do not use for clinical diagnosis.

**STORAGE AND THE TERM OF VALIDITY**

Storage Condition : 2 - 8°C  
The expiry date is specified on outer box.

Version 4

October 2016 \*

Made in Japan

**PERFORMANCE CHARACTERISTICS**
**1. Titer Assay (Samples with standard added are used.)**

Specimen	Titer (X)	Measurement Value (pg/mL)	Theoretical Value (pg/mL)	%
10% FCS added RPMI-1640	2	3,886.48	4,000.00	97.2
	4	1,996.41	2,000.00	99.8
	8	1,038.40	1,000.00	103.8
Mouse Serum (BALB/c)	4	1,256.81	2,108.80	59.6
	8	871.06	1,067.60	81.6
	16	523.64	544.01	96.3
Mouse Plasma (EDTA) (BALB/c)	4	1,598.67	2,108.80	75.8
	8	909.03	1,068.86	85.0
	16	545.95	541.57	100.8

**2. Added Recovery Assay**

Specimen	Theoretical Value (pg/mL)	Measurement Value (pg/mL)	%
10% FCS added RPMI-1640 (x2)	2,000.00	1,861.88	93.09
	1,000.00	993.26	99.33
	500.00	505.53	101.11
Mouse Serum (BALB/c) (x8)	2,045.85	1,696.85	82.94
	1,045.85	900.11	86.06
	545.85	440.39	80.68
Mouse Plasma (EDTA) (BALB/c) (x8)	2,050.35	1,663.53	81.13
	1,050.35	876.26	83.43
	550.35	440.85	80.10

**3. Intra - Assay**

Measurement Value (pg/mL)	SD value	CV value (%)	n
295.39	14.23	4.8	21
928.53	59.83	6.4	21
2,982.30	122.77	4.1	21

**4. Inter - Assay**

Measurement Value (pg/mL)	SD value	CV value (%)	n
293.49	29.86	10.2	14
864.07	44.25	5.1	14
2,916.50	80.71	2.8	14

**5. Specificity**

Compound	Cross Reactivity
Mouse VEGF 164	100.0 %
Human VEGF 165	≤0.1 %

**6. Sensitivity**

10.98 pg/mL

The sensitivity for this kit was determined using the guidelines under the National Committee for Clinical Laboratory Standards (NCCLS) Evaluation Protocols. (National Committee for Clinical Laboratory Standards Evaluation Protocols, SC1, (1989) Villanova, PA: NCCLS.)

**PRECAUTION FOR INTENDED USE AND/OR HANDLING**

- All reagents should be stored at 2 - 8°C. All reagents shall be brought to room temperature approximately 30 minutes before use.
- "3, Standard" is lyophilized products. Be careful to open this vial.
- "7, Stop solution" is a strong acid substance. Therefore, be careful not to have