



## Specificity Test (Cross Reaction Test with $\beta$ -Klotho)

Report Made: 10th Dec 2020  
 Code No. 27998  
 Product Name: Human soluble  $\alpha$ -Klotho Assay Kit - IBL

$\alpha$ -Klotho		$\beta$ -Klotho	
pg/mL	A450	pg/mL	A450
6,000	2.391	6,000,000	0.006
3,000	1.444	3,000,000	-0.001
1,500	0.749	1,500,000	0.000
750	0.380	750,000	-0.003
375	0.185	375,000	-0.005
188	0.088	187,500	-0.002
94	0.045	93,750	-0.003

